

# Microcontroller Leak Test Apparatus 611 - E



## TECHNICAL SPECIFICATIONS

Body Material:	S.S. 316
Operation:	Fully – Automatic
Set – Up Time:	Programmable – 1 to 9999 sec
Hold Time:	Programmable – 1 to 999 sec
Vacuum Level:	Programmable up to 560 mmHg
Vacuum Read:	Digital Display
Display:	20 x 4 Alphanumeric LCD
Desiccators (Optional):	150 mm / 200 mm / 300 mm Diameter
Power:	230 V ac $\pm$ 10%, 50 Hz
Physical Dimensions(H x W x D):	180 mm (max) x 320 mm x 400 mm
Weight:	9 Kg Approximately

**nics 96**

ISO 9001:2008 CERTIFIED COMPANY

PRECISE ANALYTICAL INSTRUMENTS

Manufacturing unit: Icon Automation, 25+26/18, Nilsin plot, Phase-1, GIDC – Vatva, Ahmedabad – 382 445, Gujarat State. Web-site: [www.nics96.com](http://www.nics96.com), Mail: [nics96ahm@gmail.com](mailto:nics96ahm@gmail.com), [iconautomation@yahoo.com](mailto:iconautomation@yahoo.com)