Microcontroller pH Meter 217 - E



TECHNICAL SPECIFICATIONS

0 to 14 pH
0 to 1999 <u>+</u> 1 mv
0 to 100 Degree Centigrade
0.001 pH, 1 mv, 0.1 Degree
0.01 pH <u>+</u> 1 Digit, 1 mv <u>+</u> 1 Digit, 0.2 Degree
5 point calibration Automatic control
0 to 100 Degree – Auto as well as manual mode
20 x 4 Alphanumeric LCD Display
Dot matrix Parallel interface port provided
230 Vac ± 10%, 50 Hz
80mm (max) x 238mm x 205 mm
1.5 Kg Approximately
Electrode, Stand-Clamp-Rod, Dust Cover



ISO 9001:2008 CERTIFIED COMPANY

PRECISE ANALYTICAL INSTRUMENTS

Manufacturing unit: Icon Automation, 25+26/18, Nilsin plot, Phase-1, GIDC – Vatva, Ahmedabad – 382 445, Gujarat State. Web-site: www.nics96.com, Mail: nics96ahm@gmail.com, iconautomation@yahoo.com